


<b>Brand Name of Product</b>	AquaTest™
<b>Generic Name of Product</b>	Water Test
<b>Product Code Number(s)</b>	AT101
<b>Purpose of Product</b>	To test water quality critical to proper cleaning, including pH, hardness and alkalinity.
<b>Range of Applications for Product</b>	Water used in the cleaning process
<b>Key Specifications of Product</b>	<ul style="list-style-type: none"> <li>• 3- in-1 water quality test strips. <ul style="list-style-type: none"> <li>○ pH (2 – 12)</li> <li>○ Alkalinity (0 – 500ppm CaCO<sub>3</sub>)</li> <li>○ Hardness (0 – 1000 ppm CaCO<sub>3</sub>)</li> </ul> </li> <li>• Result within 30 seconds.</li> <li>• Easy to interpret colormetric result.</li> </ul>

### Shipping & Storage

<b>Shipping Conditions &amp; Requirements</b>	
<b>Storage Conditions</b>	
<b>Packaging Conditions</b>	
<b>Shelf Life</b>	18 months from date of manufacturer.

### Instructions for Using Product

<b>Description of Use (s)</b>	To test water quality used in the cleaning of surgical instruments.
<b>Preparation for Use</b>	
<b>Diagrams (drawings, pictures)</b>	
<b>Steps for Use of Product</b>	<ul style="list-style-type: none"> <li>• Collect water from the cold water source into a clean container: <ol style="list-style-type: none"> <li>1. Dip entire strip into water for 5 seconds, then remove.</li> <li>2. Shake once briskly to remove excess water from test strip.</li> <li>3. Wait 20 seconds.</li> <li>4. Compare each pad to the corresponding color chart to determine the value for pH, alkalinity and hardness.</li> </ol> </li> <li>• Collect water from the hot water source into a clean container: <ol style="list-style-type: none"> <li>1. Dip entire strip into water for 5 seconds, then remove.</li> <li>2. Shake once briskly to remove excess water from test strip.</li> <li>3. Wait 20 seconds.</li> <li>4. Compare each pad to the corresponding color chart to determine the value for pH, alkalinity and hardness.</li> </ol> </li> </ul>
<b>Interpretation of Results</b>	<ol style="list-style-type: none"> <li>1. Color for each pad corresponds to a value which is indicated on the chart.</li> <li>2. Record results on the log sheet.</li> <li>3. Report any deviations from expected values.</li> </ol>
<b>Contraindications of Test Results</b>	
<b>Documentation</b>	
<b>Special Warnings and Cautions</b>	
<b>Disposal</b>	

### Reprocessing Instructions

<b>Point of Use</b>	
<b>Preparation for Decontamination</b>	
<b>Disassembly Instructions</b>	
<b>Cleaning – Manual</b>	
<b>Cleaning – Automated</b>	
<b>Disinfection</b>	
<b>Drying</b>	
<b>Maintenance, Inspection, and Testing</b>	
<b>Reassembly Instructions</b>	

<b>Packaging</b>	
<b>Sterilization</b>	
<b>Storage</b>	
<b>Additional Information</b>	

<b>Related Healthmark Products</b>	
<b>Other Product Support Documents</b>	
<b>Reference Documents</b>	ProFormance™ Monitoring Products Brochure, ProFormance™ Price List
<b>Customer Service Contact</b>	Healthmark Industries Company, Inc 33671 Doreka Fraser, MI 48026 1-586-774-7600 <a href="mailto:healthmark@hmark.com">healthmark@hmark.com</a> <a href="http://hmark.com">hmark.com</a>

2018-07-12 msmith