CartWashCheck™ Test Log

Make/ Model of the Medical Cart Washer
_____

Date Tested	Testers Initials	Item Tested	Cycle Selection	Temperature Result	Water Exposure Result	Action Taken	Other
1/1/11	SMK	Cart Cycle	Case Cart	120°F	White	Fail, did not reach correct Temp.	Call Repair Company to test temperature
1/3/11	SMK	Cart Cycle	Case Cart	180°F	White	Pass	

Observations:_____
_____
_____
_____

*Record all results: correct result should be white background and a minimum of 120ºF; on CartWashCheck™. Please note that 120°F might not be the temperatures required at each facility consult your management and manufacture of the equipment for correct temperature for the cycle selected. The special hydrophilic ink will change color (from black/dark blue to white) only if moistened by water. This demonstrates that water is reaching the area tested and indicates proper mechanical action by the washer. White Means Water Reached the Test. The square needs to be completely white no black/dark blue color left on the square.