



|  |  |
|--|--|
| <b>Brand Name of Product</b>             | TEMP-HL  |
| <b>Generic Name of Product</b>           | Temperature Monitoring   |
| <b>Product Code Number(s)</b>            | TEMP-HL  |
| <b>Purpose of Product</b>                | Irreversible, disposable thermometer has the ability to indicate both high and low temperatures to ensure that reprocessed items stay within a desired temperature range during transportation of goods. |
| <b>Range of Applications for Product</b> | Instrument sets and other equipment that has been reprocessed and requires temperature monitoring.   |
| <b>Key specifications of product</b>     | Irreversible Thermometer   |

**Shipping & Storage**

|   |  |
|---|--|
| <b>Shipping Conditions &amp; Requirements</b> |  |
| <b>Storage Conditions</b>                     |  |
| <b>Packaging Conditions</b>                   |  |
| <b>Shelf Life</b>                             |  |

**Instructions for Using Product**

|  |   |
|--|---|
| <b>Description of Use (s)</b>            | Thermometer indicates both high and low temperatures to ensure that reprocessed items stay within a desired temperature range during transportation of goods.   |
| <b>Preparation for Use</b>               | Be sure that instrument sets and equipment are reprocessed and dry before applying.   |
| <b>Diagrams (drawings, pictures):</b>    |   |
| <b>Steps for Use of Product</b>          | <ol style="list-style-type: none"> <li>1. Remove adhesive and place on the desired location where temperature is being monitored.</li> <li>2. When the temperature gets to 0°C, a distinct clear to red color change occurs. Conversely, when the temperature exceeds 29°C, the color changes from white to black.</li> </ol> |
| <b>Interpretation of Results</b>         |   |
| <b>Contraindications of Test Results</b> |   |
| <b>Documentation</b>                     |   |
| <b>Special Warnings and Cautions</b>     |   |
| <b>Disposal</b>                          | <ul style="list-style-type: none"> <li>• If adhesive residue remains on the surface, may be removed with acetone or isopropyl alcohol solvent.</li> <li>• May be disposed of in regular trash.</li> </ul>   |

**Reprocessing Instructions**

|  |  |
|--|--|
| <b>Point of use:</b>                         |  |
| <b>Preparation for decontamination:</b>      |  |
| <b>Disassembly Instructions:</b>             |  |
| <b>Cleaning – Manual:</b>                    |  |
| <b>Cleaning – Automated:</b>                 |  |
| <b>Disinfection:</b>                         |  |
| <b>Drying:</b>                               |  |
| <b>Maintenance, inspection, and testing:</b> |  |
| <b>Reassembly Instructions:</b>              |  |
| <b>Packaging:</b>                            |  |
| <b>Sterilization:</b>                        |  |
| <b>Storage:</b>                              |  |
| <b>Additional Information:</b>               |  |

|  |  |
|--|--|
| <b>Related Healthmark Products</b>     |  |
| <b>Other Product Support Documents</b> | Proformance Brochure, Proformance Price List       |
| <b>Reference Documents</b>             |  |
| <b>Customer Service contact:</b>       | Healthmark Industries Company, Inc<br>33671 Doreka |

|  |   |
|--|---|
|  | Fraser, MI 48026<br>1-586-774-7600<br><a href="mailto:healthmark@hmark.com">healthmark@hmark.com</a><br>hmark.com |
|--|---|

2016-02-24 Matt Smith