



Instructions for Use: Bunt Instrument Holder

Brand Name of Product	Bunt Instrument Holder
Generic Name of Product	Stringer
Product Code Number(s)	10-1770
Intended Use	This product is used to keep instruments together during cleaning and sterilization.
Range of Applications for Product	N/A
Key Specifications of Product	Manufactured of stainless steel

Shipping & Storage	
Shipping Conditions & Requirements	N/A
Storage Conditions	N/A
Packaging Contents	N/A
Shelf Life	N/A

Instructions for Using Product	
Description of Use(s)	Keeps instruments together during cleaning and sterilization.
Preparation for Use	N/A
Diagrams (drawings, pictures):	
Steps for Use of Product	<ol style="list-style-type: none"> 1. Press down on hinge to open the instrument holder pin. 2. The loop end will spring open to begin stringing instruments. 3. Place each instrument securely onto the instrument holder. 4. Continue until all instruments are loaded onto the instrument holder. 5. Lock the stringer by once again pressing down. The looped end will lock securely into place.
Interpretation of Results	N/A
Contraindications of Test Results	N/A
Documentation	N/A
Special Warnings and Cautions	N/A
Disposal	N/A

Reprocessing Instructions	
Point of Use	N/A
Preparation for Decontamination	N/A
Disassembly Instructions	N/A
Cleaning – Manual	<ul style="list-style-type: none"> • Before initial use of the instrument holder, cleaning is required.

	<ul style="list-style-type: none"> ○ Manually wash instrument holder with a neutral pH detergent and warm water. Expand and collapse the adjustable assembly area and actuate locking mechanisms during the washing process. ○ Place the instrument holder in ultrasonic cleaner for effective cleaning. ○ Place instrument holder in automated washer system for effective cleaning. ● After initial use, the instrument holder is adequately cleaned during the cleaning of the instruments.
Cleaning – Automated	<ol style="list-style-type: none"> 1. See instructions under “Steps for Use of Product” for directions on how to string instruments. 2. When using for the cleaning of instruments in an automated washer, open instrument holder as wide as possible to maximize exposure of the instrument to the cleaning solution. Consult the instrument manufacturer’s IFU for specific instructions.
Disinfection	N/A
Drying	N/A
Maintenance, Inspection, and Testing	Apply instrument milk to sliding cross bar mechanism as needed to ensure smooth action.
Reassembly Instructions	N/A
Packaging	N/A
Sterilization	<ol style="list-style-type: none"> 1. See instructions under “Steps for Use of Product” for directions on how to string instruments. 2. When using for sterilization of instruments, open to the optimal width for use in the instrument set. Consult the instrument manufacturer’s IFU for specific instructions. 3. Compatible with steam and EO sterilization. 4. Follow the sterilization instructions for the instruments that are being sterilized. 5. Maximum steam sterilization at 135°C for 30 minutes.
Storage	N/A
Additional Information	N/A

Related Healthmark Products	Stringers
Other Product Support Documents	ProSys™ Brochure, ProSys™ Price List
Reference Documents	N/A
Customer Service Contact	Healthmark Industries Company, Inc. 18600 Mayln Blvd. Fraser, MI 48026 1-586-774-7600 healthmark@hmark.com hmark.com