

Lint-Free Scope Wipes

Manual cleaning of endoscopes are necessary with low linting wipes immediately after removing the endoscope from the patient, and prior to automated or manual disinfection.

Use Lint-Free Wipes to clean and dry the endoscopes to help insure proper reprocessing.



Code No.	QTY
CC555	200


Endo: 1/2012



healthmark
INDUSTRIES CO.
health care products

800-521-6224
www.hmark.com

Endoscopy Products



Healthmark Industries

Call 800-521-6224 or visit HMARK.COM

ChannelCheck™


ChannelCheck™ is the first product capable of testing virtually any lumened instrument for residual organic soils, including flexible endoscopes, no matter the channel size.



Code No.	QTY
UCC-101	100

Disinfection Labels

Keep track of when your products are disinfected with specific labels.




Code No.	Description	Color
DC-100214-71	Disinfected Label	Blue

Hang Time Kit™

The first step in managing scope storage is to know when they were last reprocessed. For that purpose, the kit includes labels to track by day and month.


What about a scope that has been stored for more than a couple a days? The kit also includes the HTK Test. This simple and rapid test checks for gram negative bacteria that might have grown out during storage.



Code No.	Description	Qty
403225 HANG TIME	Label with Month and Dates	500
403115	White Tray Label	500
403225	Yellow Tray Label	500
404225-351-GI	Self Looping Endoscope Reprocessing 13" Label	500
FV-1041-ES	Gram Negative Bacteria Test, 1ml Sterile Pipette	50

Flexicheck™

This three part kit simulates a flexible endoscope channel and is designed to challenge the cleaning efficiency of endoscope washers with channel irrigation apparatus. The kit includes a clear flexible tube, attached to a stainless steel lumen device.



Code No.	Description	QTY
FLEXCHK-101	Flexi-Check Flexible Scope Washer	25
FLEXCHK-102	Flexi-Check Holder	1

Clean Cynch-Lok

These locks can be pulled to varied degrees of tightness. They are ideal when a tight fitting seal is required.



Code No.	Description
5030 White "Clean"	White Pulltight - "Clean"

Endoscope Brushes

Long enough, small enough and flexible enough to clean your flexible endoscopes.



Code No	Description	Dia (mm)	Lgth (mm)	Trim (mm)
00711610	Control Head/Valve Cleaning Brush	1.0	4	10
007-601	Standard Colonoscope/ Duodenoscope	1.7	230	6
007-603	Jumbo Colonoscope/ Duodenoscope	1.7	230	9
CC-110	Clear Channel Brush	1.2	65	
CC-250	Clear Channel Brush (2.8 - 3.7 mm)	2.8	250	
ENDO-BRK-01	Endoscope Brush Kit - includes 1 each: 00711601, 00711610, 3770, CC-250			

Endo Parts Bag

The Endo Parts Bag allows you to keep all valves for each scope together.

Code No.	QTY
EBP-46	10



EndoCheck™

Flexible endoscopes are difficult to clean. More difficult still is verifying that the cleaning has been adequate. Now there is a test for checking the cleanliness of the biopsy channel of the flexible endoscope.



Code No.	Description
EDH-110	EndoCheck: Hemoglobin Residue for 1.1mm x 200cm lumen (12 per box)
EDH-200	EndoCheck: Hemoglobin Residue for 2mm x 230cm lumen (12 per box)
EDH-270	EndoCheck: Hemoglobin Residue for 2.7mm x 230cm lumen (12 per box)
EDH-350	EndoCheck: Hemoglobin Residue for 3.5mm x 230cm lumen (12 per box)
EDH-470	EndoCheck: Hemoglobin Residue for 4.7mm x 230cm lumen (12 per box)
EDP-110	EndoCheck: Hemoglobin Residue for 1.1mm x 200cm lumen (12 per box)
EDP-200P	EndoCheck: Hemoglobin Residue for 2mm x 230cm lumen (12 per box)
EDP-270	EndoCheck: Hemoglobin Residue for 2.7mm x 230cm lumen (12 per box)
EDP-350	EndoCheck: Hemoglobin Residue for 3.5mm x 230cm lumen (12 per box)
EDP-470	EndoCheck: Hemoglobin Residue for 4.7mm x 230cm lumen (12 per box)

HumiPak®

Flexible endoscope manufacturers, AAMI, AORN, the Multi-Society and others generally recommend that decontamination of scopes begin within 30 minutes of use so that organic soils do not dry. But often this is not possible. The solution to extended holding time of soiled scopes? The patented Humipak!

Testing has demonstrated that bodily fluids will remain moist for up to 3 days, depending upon the exact composition of the residue. Keep scopes moist and ready for effective cleaning with the Humipak!



Code No.	Description	Width	Length
HPSS2035	Double Din Size	26.5	25.5