



Instructions for Use: Silicone Cable & Instrument Ties

Brand Name of Product	Silicone Cable & Instrument Ties
Generic Name of Product	Bundling Tool
Product Code Number(s)	09039, 09143, 09144, 09145, 09146, 09147, 09196, 09708, 09139, 09160, 09161, 09163, 09164, 09288, 09709, 09231, 09235, 09236, 09237, 09238, 09369, 09370, 09239, 09240, 09241, 09242, 09371, 09372, 09285
Intended Use	Ties are used to simplify the process of bundling items.
Range of Applications for Product	N/A
Key Specifications of Product	<ul style="list-style-type: none"> Available in a variety of colors

Shipping & Storage	
Shipping Conditions & Requirements	N/A
Storage Conditions	N/A
Packaging Contents	N/A
Shelf Life	N/A

Instructions for Using Product	
Description of Use(s)	N/A
Preparation for Use	N/A
Diagrams (drawings, pictures)	N/A
Steps for Use of Product	Simply wrap the items pull the end through the locking mechanism or tie for desired tightness.
Interpretation of Results	N/A
Contraindications of Test Results	N/A
Documentation	N/A
Special Warnings and Cautions	N/A
Disposal	Single-use; dispose of in accordance with your facility's guidelines.

Reprocessing Instructions	
Point of Use	N/A
Preparation for Decontamination	N/A
Disassembly Instructions	N/A
Cleaning – Manual	N/A
Cleaning – Automated	N/A
Disinfection	N/A
Drying	N/A
Maintenance, Inspection, and Testing	N/A
Reassembly Instructions	N/A
Packaging	N/A
Sterilization	<ul style="list-style-type: none"> Steam Sterilization- 132°C for 4 minutes or 135°C for 3 minutes. Materially compatible with EO and VH202 sterilization.
Storage	N/A
Additional Information	N/A

Related Healthmark Products	N/A
Other Product Support Documents	ProSys™ Brochure, ProSys™ Price List
Reference Documents	N/A
Customer Service Contact	Healthmark Industries Company, Inc. 18600 Malyn Blvd. Fraser, MI 48026 1-586-774-7600 healthmark@hmark.com hmark.com

