

Instructions for Use: Spring Locking Instrument Stringers

Brand Name of Product	Spring Locking Instrument Stringers	
Generic Name of Product	Instrument stringers	
Product Code Number(s)	USPRS-10, USPRS-12, USPRS-14, USPRS-16, USPRS-6, USPRS-8	
Purpose of Product	This product is used to expand instruments to better expose hard to reach areas during	
_	cleaning and sterilization.	
Range of Applications for Product		
Key specifications of product	Manufactured of stainless steel	

Shipping & Storage		
Shipping Conditions & Requirements		
Storage Conditions		
Shelf Life		

Instructions for Using Product			
Description of Use (s)	Keeps instruments maximally exposed for hard to reach areas during cleaning and		
	sterilization.		
Preparation for Use			
Diagrams (drawings, pictures):			
	1	2	3
Steps for Use of Product	 Unlock the stringer by gently pulling the smaller center piece of the spring out from inside the larger piece. Once the latch is unlocked, extend both sides of the latch so they are completely extended. String ring handled instruments in set assembly order onto stringer by placing one ring on each of the stringer arms. Continue until all instruments are loaded onto stringer. Lock latch by closing the latch into a secure locking position. The smaller end of the spring will insert inside the larger spring. 		
Interpretation of Results			
Contraindications of Test Results			
Documentation			
Special Warnings and Cautions			
Disposal			

Reprocessing Instructions		
Point of use:		
Preparation for decontamination:		
Disassembly Instructions:		
Cleaning – Manual:	Before initial use of the stringer, cleaning is required.	
	 Manually wash stringer(s) with a neutral ph detergent and warm water. 	
Cleaning - Automated:		
Disinfection:		
Drying:		
Maintenance, inspection, and testing:		
Reassembly Instructions:		
Packaging:		
Sterilization:	Stringers are materially compatible with steam and EO sterilization.	

Storage:	
Additional Information:	

Related Healthmark Products	Stringers	
Other Product Support Documents	Prosys Price List, Prosys Brochure	
Reference Documents		
Customer Service contact:	Healthmark Industries Company, Inc	
	33671 Doreka	
	Fraser, MI 48026	
	1-586-774-7600	
	healthmark@hmark.com	
	hmark.com	

2014-05-02 msmith