
















Instructions for Use: Weinstein Instrument Holder

Brand Name of Product	Weinstein Instrument Holder
Generic Name of Product	Instrument Stringers
Product Code Number(s)	HE1-0006-W, HE1-0008-W, HE1-0010-W, HE1-0012-W, HE1-0014-W, HE1-0016-W, HE1-0018-W, 10-1805, TBIS-8
Purpose of Product	This product is used to expand instruments to better expose hard-to-reach areas during cleaning and sterilization.
Range of Applications for Product	N/A
Key Specifications of Product	<ul style="list-style-type: none"> Manufactured of stainless steel. Available in a variety of widths and lengths.

Shipping & Storage	
Shipping Conditions & Requirements	N/A
Storage Conditions	N/A
Packaging Contents	N/A
Shelf Life	N/A

Instructions for Using Product													
Description of Use(s)	Keeps instruments maximally exposed for hard-to-reach areas during cleaning and sterilization.												
Preparation for Use	N/A												
Diagrams (drawings, pictures)	<table border="1"> <tr> <td></td> <td></td> <td></td> </tr> <tr> <td>1</td> <td>2</td> <td>3</td> </tr> <tr> <td></td> <td></td> <td></td> </tr> <tr> <td>4</td> <td>5</td> <td></td> </tr> </table>				1	2	3				4	5	
													
1	2	3											
													
4	5												

Steps for Use of Product	<ol style="list-style-type: none"> 1. Place one hand at each end of the stringer. 2. Gently squeeze the stringer arms inwards and pull to separate the stringer from the top. 3. String handled instruments (in set assembly order) onto the stringer by placing one ring on each of the stringer arms. <ol style="list-style-type: none"> a. Continue until all instruments are loaded onto the stringer. b. Lock the stringer by reattaching the top. (NOTE: You may need to gently squeeze one of the arms inward to thread into the crescent end.)
Interpretation of Results	N/A
Contraindications of Test Results	N/A
Documentation	N/A
Special Warnings and Cautions	N/A
Disposal	N/A

Reprocessing Instructions	
Point of Use	N/A
Preparation for Decontamination	N/A
Disassembly Instructions	N/A
Cleaning – Manual	<ul style="list-style-type: none"> ● Before initial use of the stringer, cleaning is required before initial use of the stringer. <ul style="list-style-type: none"> ○ Manually wash using neutral pH detergent and warm water. ○ Place the stringer(s) in ultrasonic cleaner for effective cleaning. ○ Next place stringer(s) in automated washer system for effective cleaning. ● After initial use, the stringer is adequately cleaned during the cleaning of the instruments.
Cleaning – Automated	See instructions under “Steps for Use of Product” for directions on how to string instruments.
Disinfection	N/A
Drying	N/A
Maintenance, Inspection, and Testing	N/A
Reassembly Instructions	N/A
Packaging	N/A
Sterilization	<ol style="list-style-type: none"> 1. See instructions under “Steps for Use of Product” for directions on how to string instruments. 2. When using for sterilization of instruments: <ol style="list-style-type: none"> a. Open to the optimal width for use in the instrument set b. Consult the instrument manufacturer’s IFU for specific instructions. 3. Stringers are materially compatible with steam and EO sterilization. 4. Maximum steam sterilization at 135 °C (275 °F) for 30 minutes.
Storage	N/A
Additional Information	N/A

Related Healthmark Products	N/A
Other Product Support Documents	ProSys™ Brochure, ProSys™ Price List
Reference Documents	N/A
Customer Service Contact	Healthmark Industries Company, Inc. 18600 Malyn BLvd. Fraser, MI 48026 1-586-774-7600 healthmark@hmark.com hmark.com