



Instructions for Use: Shark Tooth Heat Mitt

Brand Name of Product	Shark Tooth Heat Mitt
Generic Name of Product	Silicone Heat Mitt
Product Code Number(s)	SLG-001 SLG-002
Purpose of Product	The shark tooth ridges provide a safe and secure grip of hot and wet items.
Range of Applications for Product	For handling hot instruments, trays and equipment.
Key specifications of product	<ul style="list-style-type: none"> • Made of heat deflecting silicone that is sterilizable, washable, and resistant to temperatures from -40°F to 446°F. • The shark tooth ridges provide a safe and secure grip of hot and wet items. • Sizes: <ul style="list-style-type: none"> ○ SLG-001 Blue 14” length ○ SLG-002 Black 17” length

Shipping & Storage	
Shipping Conditions & Requirements	
Storage Conditions	Room temperature
Packaging Conditions	1 each Shark Tooth Heat Mitt per package
Shelf Life	

Instructions for Using Product	
Description of Use (s)	
Preparation for Use	
Diagrams (drawings, pictures):	
Steps for Use of Product	<ol style="list-style-type: none"> 1. Insert hand completely into the mitt. 2. Mitt can be worn on either right or left hand.
Interpretation of Results	
Contraindications of Test Results	
Documentation	
Special Warnings and Cautions	
Disposal	

Reprocessing Instructions	
Point of use:	
Preparation for decontamination:	
Disassembly Instructions:	
Cleaning – Manual:	<ol style="list-style-type: none"> 1. The mitt may be cleaned manually utilizing a neutral PH detergent and a soft bristle nylon brush such as Healthmark brushes 869903N, GDS-1 or GDS-2 and or a soft lint free cleaning cloth. 2. After cleaning, the mitt should be copiously rinsed with tap water. 3. The mitt may be air dried or dried with a lint free cloth or towel.
Cleaning – Automated:	
Disinfection:	<ol style="list-style-type: none"> 1. If desired, after cleaning and drying, the mitt may be disinfected utilizing a liquid chemical disinfectant, which is compatible with silicone, in accordance with the chemical manufacturers IFUs. 2. Following disinfection, the mitt should be copiously rinsed with tap water to remove any chemical residuals.
Drying:	The mitt may be air dried or dried with a lint free clean cloth or towel.
Maintenance, inspection, and testing:	
Reassembly Instructions:	

Packaging:	
Sterilization:	If needed, the mitt may be processed through a pre vac steam sterilizer in accordance with the healthcare facility's procedures and conventional cycles at temperatures not to exceed 446 °F.
Storage:	
Additional Information:	

Related Healthmark Products	Insulated Gauntlet, Seamless Autoclave Gloves, Terry Cloth Autoclave Gloves.
Other Product Support Documents	Prosys Brochure, Prosys Price List
Reference Documents	
Customer Service contact:	Healthmark Industries Company, Inc 33671 Doreka Fraser, MI 48026 1-586-774-7600 healthmark@hmark.com hmark.com

2012-07-17 msmith