



Instructions for Use: Silicone Heat Glove

Brand Name of Product	Silicone Heat Glove
Generic Name of Product	Heat Glove
Product Code Number(s)	09484, 09368
Intended Use	For the safe handling of hot trays and racks from autoclaves
Range of Applications for Product	Tray, racks and items which are hot that are being removed from the autoclave
Key specifications of product	<ul style="list-style-type: none"> Made of Silicone Protects from surface temperatures ranging from -40°F to +464°F

Shipping & Storage	
Shipping Conditions & Requirements	
Storage Conditions	Stored at room temperature. Do not expose to direct sunlight.
Packaging Conditions	
Shelf Life	

Instructions for Using Product	
Description of Use (s)	
Preparation for Use	
Diagrams (drawings, pictures):	
Steps for Use of Product	<ol style="list-style-type: none"> Insert hands completely into the glove. Be sure that the glove is completely covering your hand for maximum protection.
Interpretation of Results	
Contraindications of Test Results	
Documentation	
Special Warnings and Cautions	
Disposal	

Reprocessing Instructions	
Point of use:	
Preparation for decontamination:	
Disassembly Instructions:	
Cleaning – Manual:	Manual or mechanical reprocessing without ultrasound treatment
Cleaning – Automated:	Acidic / neutral / alkaline >pH 10 with or without tenside additives, chemically at max. 60° C / 140°F or with VE water, thermally, at max. 93°C / 199°F
Disinfection:	
Drying:	
Maintenance, inspection, and testing:	
Reassembly Instructions:	
Packaging:	
Sterilization:	Compatible with steam sterilization
Storage:	
Additional Information:	

Related Healthmark Products	Seamless Autoclave Gloves, Terry Cloth Autoclave Gloves, Shark Tooth Heat Mitts
Other Product Support Documents	Prosys Brochure, Prosys Price List
Reference Documents	
Customer Service contact:	Healthmark Industries Company, Inc 33671 Doreka Fraser, MI 48026 1-586-774-7600 healthmark@hmark.com hmark.com