


| | |
|--|--|
| Brand Name of Product | Made in USA LED Magnifier |
| Generic Name of Product | Magnifying Lamp |
| Product Code Number(s) | 42400-4 RD, 82400-4BL, 11426 |
| Intended Use | Designed for professional optical inspection of items and equipment. |
| Range of Applications for Product | Any products that require optical inspection after washing and before being sterilized. |
| Key Specifications of Product | <ul style="list-style-type: none"> • 7" X 5¼" RECTANGULAR LED MAGNIFIER - 43" REACH - TABLE EDGE CLAMP (Blue) • 6" ROUND LED MAGNIFIER - 43" REACH -SCREW DOWN BASE – Red • Assembled with low voltage (15-24v) power supplies. Shuts off every 12 hours to curb power supplies. • 3" wide with corner holes mount (Sold Separately) • 2x magnification |

Shipping & Storage

| | |
|---|--|
| Shipping Conditions & Requirements | |
| Storage Conditions | |
| Packaging Conditions | |
| Shelf Life | |

Instructions for Using Product

| | |
|--|---|
| Description of Use (s) | Designed for visual inspection for items and equipment. |
| Preparation for Use | |
| Diagrams (drawings, pictures) |  |
| Steps for Use of Product | <ol style="list-style-type: none"> 1) Attach the lamp to the desired working surface. If using the optional bracket to mount please note: <ol style="list-style-type: none"> a. Mounting to typical wall with 2"x 4" studs, there will be a thin side of the stud, approximately 1.75" to anchor to. The holes will not allow for all four to be anchored on a stud this size. It is also NOT advisable to anchor this directly to dry wall as it would probably fall off. 2) Position the lamp in the desired position and turn on using the on-off switch. |
| Interpretation of Results | |
| Contraindications of Test Results | |
| Documentation | |
| Special Warnings and Cautions | |
| Disposal | |

Reprocessing Instructions

| | |
|---------------------|--|
| Point of Use | |
|---------------------|--|

| | |
|---|---|
| Preparation for Decontamination | |
| Disassembly Instructions | |
| Cleaning – Manual | <ul style="list-style-type: none"> ○ Clean with a soft cloth moistened in warm, soapy water or use a lens cleaner. ○ Dry with clean soft cloth (never use abrasive cleaners). |
| Cleaning – Automated | |
| Disinfection | |
| Drying | |
| Maintenance, Inspection, and Testing | |
| Reassembly Instructions | |
| Packaging | |
| Sterilization | |
| Storage | |
| Additional Information | |

| | |
|--|--|
| Related Healthmark Products | |
| Other Product Support Documents | ProSys™ Brochure, ProSys™ Price List |
| Reference Documents | |
| Customer Service Contact | Healthmark Industries Company, Inc 33671 Doreka Fraser, MI 48026 1-586-774-7600 healthmark@hmark.com hmark.com |

2018-08-13 msmith