



Instructions for Use: Magic Touch Magnifier

Brand Name of Product	Magic Touch Magnifier
Generic Name of Product	Magnifier
Product Code Number(s)	RNDMAG-1, RNDMAGSTND-1
Purpose of Product	A must-have tool with a small size that makes it convenient to carry at all times. Designed to allow for easy and thorough inspection of surgical instruments.
Range of Applications for Product	Surgical instruments that need up close visual inspection for possible damage or residual soiling after reprocessing.
Key specifications of product	<ul style="list-style-type: none"> • 3x magnification • Lightweight .45 lbs and durable design • 3 LED lights • Touch activated and on/off switch • Microfiber lens cloth

Shipping & Storage	
Shipping Conditions & Requirements	
Storage Conditions	Always store loupe in included protective case when not in use to avoid damaging the lens.
Packaging Conditions	
Shelf Life	

Instructions for Using Product	
Description of Use (s)	Close-up visual inspection of instruments after reprocessing
Preparation for Use	<p>Batteries come installed. When changing batteries:</p> <p>Turn item upside down and locate battery compartment. (Please note there are two battery compartments. The compartment closest to the light switch is the one that powers the item. The other battery compartment is used to store batteries that are not in use.) Pull battery door up and fold door away from the battery compartment (see Fig 1). Insert two fresh CR2016 coin cell batteries into battery compartment. Make sure that the positive side of the battery is facing outward (see Fig 2). Close battery door securely when finished.</p>
Diagrams (drawings, pictures):	
Steps for Use of Product	<p>1) Select the LED light setting you would like to use (see Fig. 3).</p> <ol style="list-style-type: none"> For constant light, position the switch in the middle position (marked ON). To use the “touch activated” setting, position the switch to the position marked “AUTO”.

	<p>i. With the switch on AUTO, you simply need to touch the silver ring on top of the loupe that surrounds the lens (as shown in Fig. 4) to activate the light.</p> <p>2) Once you have selected your LED light setting, place the loupe on top of the subject you would like to magnify. Focus the magnifier by rotating the top portion of the loupe until the viewing subject is sharp and clear (see Fig. 5).</p>
Interpretation of Results	
Contraindications of Test Results	
Documentation	
Special Warnings and Cautions	<ul style="list-style-type: none"> • Do not leave batteries installed during extended periods of disuse. • Do not mix new and old batteries. Do not mix alkaline, standard (carbon-zinc) or rechargeable (nickel-cadmium) batteries. • Fire Hazard. Keep magnifier away from direct sunlight to prevent accidental fire. • Do not stare directly at LED lights as this might damage your eyes.
Disposal	

Reprocessing Instructions	
Point of use:	
Preparation for decontamination:	
Disassembly Instructions:	
Cleaning – Manual:	<ul style="list-style-type: none"> • Clean lens with a soft, clean microfiber cloth rubbing gently in a circular motion. • Moisten the lens cloth with water if necessary. • Do not use harsh chemicals, solvents or abrasives as they may damage the lens.
Cleaning – Automated:	
Disinfection:	
Drying:	
Maintenance, inspection, and testing:	
Reassembly Instructions:	
Packaging:	
Sterilization:	
Storage:	
Additional Information:	

Related Healthmark Products	
Other Product Support Documents	
Reference Documents	ProSys Brochure, ProSys Price List
Customer Service contact:	Healthmark Industries Company, Inc 33671 Doreka Fraser, MI 48026 1-586-774-7600 healthmark@hmark.com hmark.com

2017-06-06 Ashlynn Basile