




Instructions for Use: Lint-Free Disposable Wipes

Brand Name of Product	Lint-Free Disposable Wipes
Generic Name of Product	Instrument Cleaning Wipes
Product Code Number(s)	CC555, CC1212, CC1212B
Purpose of Product	Low linting wipes are for use immediately after surgical equipment has been used, and prior to automated or manual disinfection to ensure items are dry.
Range of Applications for Product	Any instruments that requires immediate removal of soils after procedure.
Key specifications of product	<ul style="list-style-type: none"> • Low linting blended poly-cellulose material • Single-use

Shipping & Storage

Shipping Conditions & Requirements	
Storage Conditions	
Packaging Conditions	
Shelf Life	

Instructions for Using Product

Description of Use (s)	To allow for safe cleaning of sensitive instruments and fixtures, as well as all general surfaces.
Preparation for Use	<ol style="list-style-type: none"> 1. Be sure to consult and follow closely the device manufacturer’s IFU. 2. If utilizing a cleaning agent, be sure to closely follow the detergent manufacturer’s IFU for proper preparation of the cleaning solution.
Diagrams (drawings, pictures):	
Steps for Use of Product	<ul style="list-style-type: none"> • Moisten with cleaning solution (If applicable) • Wipe down instrument to remove soils and gross debris • Dispose after use on a single instrument
Interpretation of Results	
Contraindications of Test Results	
Documentation	
Special Warnings and Cautions	<ul style="list-style-type: none"> • Be sure to consult the IFU of the device manufacturer to be sure full compliance with their instructions. • Pay special attention to area(s) of the instrument that must be free from moisture. • Pay special attention to chemical compatibility of the medical device and areas of the medical device with the cleaning agent you select.
Disposal	Treat as biohazard and discard of in compliance with facility guidelines.

Reprocessing Instructions

Point of use:	
Preparation for decontamination:	
Disassembly Instructions:	
Cleaning – Manual:	
Cleaning – Automated:	
Disinfection:	
Drying:	
Maintenance, inspection, and testing:	
Reassembly Instructions:	
Packaging:	
Sterilization:	
Storage:	
Additional Information:	

Related Healthmark Products	DuraSponge™
Other Product Support Documents	Prosys Price Brochure, Prosys Price List
Reference Documents	
Customer Service contact:	Healthmark Industries Company, Inc 33671 Doreka Fraser, MI 48026 1-586-774-7600 healthmark@hmark.com hmark.com

2015-09-28 msmith