









**Instructions for Use: Mesh Tip Protectors**

<b>Brand Name of Product</b>	Mesh Tip Protectors
<b>Generic Name of Product</b>	Tip Protectors
<b>Product Code Number(s)</b>	CSW-03-2.0, CSW-04-4.0
<b>Intended Use:</b>	To protect delicate instruments during VHP sterilization and/or storage.
<b>Range of Applications for Product</b>	Small to large tipped instruments.
<b>Key specifications of product</b>	<ul style="list-style-type: none"> <li>• Compatible with vaporized hydrogen peroxide including Sterrad and V-Pro systems.</li> <li>• Mesh polyethylene material to promote sterilant circulation.</li> <li>• Available in closed or open ended styles (only use open ended design for flexible endoscopes).</li> <li>• Single-use</li> </ul>

<b>Shipping &amp; Storage</b>	
<b>Shipping Conditions &amp; Requirements</b>	
<b>Storage Conditions</b>	
<b>Packaging Conditions</b>	
<b>Shelf Life</b>	

<b>Instructions for Using Product</b>					
<b>Description of Use (s)</b>	To protect delicate instruments during VHP sterilization and storage.				
<b>Preparation for Use</b>					
<b>Diagrams (drawings, pictures):</b>	<table border="1" style="width: 100%;"> <tr> <td style="text-align: center;">  </td> <td style="text-align: center;">  </td> </tr> <tr> <td style="text-align: center;">Closed Style Tip</td> <td style="text-align: center;">Open Style Tip</td> </tr> </table>			Closed Style Tip	Open Style Tip
					
Closed Style Tip	Open Style Tip				
<b>Steps for Use of Product</b>	<ul style="list-style-type: none"> <li>• Place protector tip on the instrument. The Mesh style tip protector will provide some flexibility to expand to the instrument size. If open style, roll mesh tip ends of the protector to provide cushion support for the exposed tip (See Open Style Tip pictured above. )</li> <li>• Package instruments in sterile packaging device (e.g., pouch, wrapped tray, closed container).</li> <li>• Place package in sterilizer.</li> <li>• Dispose of after use in compliance with facility guidelines.</li> </ul>				
<b>Interpretation of Results</b>					
<b>Contraindications of Test Results</b>					
<b>Documentation</b>					
<b>Special Warnings and Cautions</b>					
<b>Disposal</b>	Dispose of in compliance with facility guidelines.				

<b>Reprocessing Instructions</b>	
<b>Point of use:</b>	
<b>Preparation for decontamination:</b>	
<b>Disassembly Instructions:</b>	
<b>Cleaning – Manual:</b>	
<b>Cleaning – Automated:</b>	
<b>Disinfection:</b>	
<b>Drying:</b>	

<b>Maintenance, inspection, and testing:</b>	
<b>Reassembly Instructions:</b>	
<b>Packaging:</b>	
<b>Sterilization:</b>	Compatible with vaporized hydrogen peroxide sterilization systems, including Sterrad and V-Pro
<b>Storage:</b>	
<b>Additional Information:</b>	

<b>Related Healthmark Products</b>	Tip Protectors, Vented Instrument Guards, Duo-Vented Instrument Guards, Cylindrical Style, Instrument Guards.
<b>Other Product Support Documents</b>	Mesh Style Tip Protectors-Validation for Compatibility with Hydrogen Peroxide Sterilization: <a href="http://www.healthmark.info/InstrumentCare/TipProtectors/Mesh_style_tip_protectors.pdf">http://www.healthmark.info/InstrumentCare/TipProtectors/Mesh_style_tip_protectors.pdf</a>
<b>Reference Documents</b>	Prosys Price List, Prosys Brochure
<b>Customer Service contact:</b>	Healthmark Industries Company, Inc 33671 Doreka Fraser, MI 48026 1-586-774-7600 <a href="mailto:healthmark@hmark.com">healthmark@hmark.com</a> hmark.com

2016-02-25 msmith