

ENDOSCOPE REPROCESSING QUALITY ASSURANCE RECOMMENDATIONS

	Visual Inspection*	Use of Borescope	CV Periodically**	CV Weekly or Daily**	CV with Each Scope or Daily**	Labeling for Identification^	Hazard Transport	Culturing	Reprocessing Audits	Moist Transport
AORN	✓	✓	N/A	N/A	✓	✓	✓	✓	✓	✓
SGNA	✓	N/A	✓	N/A	N/A	✓	✓	✓	✓	N/A
AAMI ST91	✓	✓	N/A	✓	N/A	✓	✓	✓	✓	N/A
Multi-Society (ASGE)	✓	N/A	✓	N/A	N/A	N/A	✓	✓	N/A	N/A
CDC - HICPAC	✓	N/A	✓	N/A	N/A	N/A	✓	✓	✓	N/A
APIC	✓	N/A	N/A	N/A	N/A	N/A	N/A	N/A	✓	N/A
OSHA	N/A	N/A	N/A	N/A	N/A	N/A	✓	N/A	N/A	N/A

✓ = has recommendations available, refer to current guidelines for exact recommendations * = Visual inspection includes lighted magnification for each scope

**CV = cleaning verification

^ = recommends a distinct visual cue affixed to scopes to see whether or not they have been reprocessed

