



Instructions for Use: Scope Transport Tray with Cover

Brand Name of Product	Scope Transport Tray
Generic Name of Product	Scope Transport Tray
Product Code Number(s)	Z-WAN-ETW-LTCH
Intended Use	For the safe collection and transport of contaminated endoscopes from the point of use to the point of reprocessing and the safe removal from the container to the decontamination area where gross contaminants can be removed.
Range of Applications for Product	For flexible endoscopes.
Key Specifications of Product	Base and cover Size- 24" x 24" OD Manufactured of Polypropylene 2 yellow latches

Shipping & Storage	
Shipping Conditions & Requirements	
Storage Conditions	
Packaging Conditions	
Shelf Life	

Instructions for Using Product	
Description of Use(s)	For safe collection, transportation and processing of contaminated endoscopes.
Preparation for Use	
Diagrams (drawings, pictures)	
Steps for Use of Product	<ol style="list-style-type: none"> 1. Transport empty scope transport tray containers to the using location (O.R., E.R., etc.). 2. Have the transport tray near using location for easy access. 3. Place contaminated endoscope in transport tray (with or without soaking solution) according to facility and the surgical instrument manufacturer's IFU. 4. Place cover on the base and push before transporting back to the decontamination room. 5. Return the scope transport tray and enclosed endoscope for decontamination. 6. Pull the latches up to remove cover and place contaminated endoscope onto flat surface to begin processing. 7. For reprocessing of the empty base tray, cover in compliance with reprocessing instructions described below. 8. Return the "clean" SST to the floor, O.R., etc., for collection and transportation.
Interpretation of Test Results	
Contraindications of Test Results	
Documentation	
Special Warnings and Cautions	<ol style="list-style-type: none"> 1. Do not nest products in the washer or in the sterilizer. 2. When washing or terminally sterilizing, be sure to arrange the trays, baskets and covers in a vertical position. 3. Do not overload the washer or the sterilizer. 4. Sterility can be maintained when the items are packaged in an FDA cleared sterile barrier system. 5. Lids with yellow latch lock easily when lid is tightly snapped on. If not snapped all the way, the latches become hard to lock down.
Disposal	

Reprocessing Instructions	
Point of Use	
Preparation for Decontamination	<ul style="list-style-type: none"> • Do not exceed recommended maximum temperature for each component of the system. See SST Material Compatibility for reprocessing parameter compatibilities for each specific model number and its component parts.

	<ul style="list-style-type: none"> Do not use any abrasive powders or metal brushes as these may cause scratching on surfaces. Reprocessing instructions apply only to an empty SST System and its component parts. These instructions are not meant or intended for processing of any reusable items.
Disassembly Instructions	
Cleaning – Manual	<ul style="list-style-type: none"> Manual cleaning is permissible but not recommended. Use equipment detergent, soft bristle brush and running water. Be sure to follow the IFU of the detergent manufacturer. Rinse excess soil from all surfaces. Apply detergent to all surfaces and use brush to clean all surfaces until they are clean and free from soil. Dry with a low linting cloth.
Cleaning – Automated	<ul style="list-style-type: none"> Machine cleaning is preferred; use a washer disinfectant meeting requirements of ISO 15883-2. Be sure to follow the IFU of the washer disinfectant and detergent manufacturers. Products should be positioned in the washer to allow maximum water penetration and drainage. <ul style="list-style-type: none"> No overlapping of items - partially covered surfaces will not be washed properly. Trays, covers and baskets should be stacked standing on the side to allow complete drainage. Products can be cleaned with alkaline, acidic and neutral detergents. For final rinse, DI or RO water is preferred. <p>The standard program to include (always refer to the washer manufacturer’s IFU):</p> <ol style="list-style-type: none"> Pre-wash with cold water (<100°F) rinse for a minimum of 2 minutes. Washing cycle with alkaline or enzymatic detergent at temperature recommended by the detergent manufacturer for a minimum of 5 minutes. Washing cycle with neutral pH or neutralizing detergent at temperature recommended by the detergent manufacturer for a minimum of 2 minutes. Final rinse cycle for a minimum of 1 minute. Thermal disinfection in compliance with the washer manufacturer recommendations for time and temperature, but not to exceed the maximum temperature tolerance for the individual SST System components. See SST Material Compatibility for temperature tolerances for each specific model number and its component parts. Drying cycle at temperature not to exceed the maximum temperature tolerance for the individual scope transport tray system components. See SST Material Compatibility for temperature tolerances for each specific model number and its component parts. Use caution when unloading machine as products will be hot. Visual inspection is required to ensure complete removal of soil. If product still shows soil, repeat program.
Disinfection	<ul style="list-style-type: none"> Disinfection may not be required depending upon the clinical use of the scope transport tray and the policy of the institution. Disinfection may be achieved by thermal means as described in the standard wash cycle above, or by use of hospital grade chemical disinfectants. If chemical disinfectants are used, it is recommended that all surfaces be thoroughly rinsed after the required exposure time. Follow the disinfectant manufacturer’s instructions for use.
Drying	Drying may be accomplished during the dry cycle of the automated washer and/or with the use of a low linting cloth.
Maintenance, Inspection, and Testing	<ul style="list-style-type: none"> Visual inspection is required to ensure complete removal of soil. If product still shows soil, repeat program. Visually inspect and assess for wear and tear and damage due to use.
Reassembly Instructions	
Packaging	
Sterilization	
Storage	

Additional Information	
Related Healthmark Products	SST Systems Products
Other Product Support Documents	SST Instrument Retrieval Products Brochure, SST Systems Products Price List, SST Material Compatibility document
Reference Documents	
Customer Service Contact	Healthmark Industries Company, Inc. 18600 Malyn Blvd. Fraser, MI 48026 1-586-774-7600 healthmark@hmark.com hmark.com

2021-04-21 Suzanne Latta