

Instructions for Use: Autoclave Tape

Brand Name of Product	Autoclave Tape		
Generic Name of Product	Autoclave Tape		
Product Code Number(s)	ATP-xxx		
Intended Use:	Steam autoclave tapes are manufactured for closing packages of items requiring steam		
	sterilization in hospitals as	nd other similar	uses.
Range of Applications for Product	Instrument wraps and other	er items that need	d to be taped closed during sterilization
Key specifications of product	Steam autoclave tapes with steam indicator are produced by using strong impregnated crepe paper, coated with an adhesive. This is principally based on natural dry rubber and tackifying synthetic resins.		
	Property Thickness: Adhesion to steel: Tensile strength: Elongation: Color: stripes	Units µ N/25 mm N/25 mm %	Typical Value 140-170 9.0 – 10.0 70 - 90 7 Natural, indicator tape printed with white

Shipping & Storage		
Shipping Conditions & Requirements		
Storage Conditions	It is recommended that the products are kept in the original, closed transport cartons and stored in dry and clean conditions protected from direct sunlight and excessive moisture	
	and heat.	
Packaging Conditions	 Steam autoclave tapes are packed in recyclable, brown hard carton cases. As an inner cover against moist and dust the steam indicator tapes are wrapped in a recyclable HD polyethylene sheet. Labelling: Each case bears a label with the necessary information/instructions of the content in accordance with ISO 11140-1:2005. 	
Shelf Life		

Instructions for Using Product			
Description of Use (s)	Use to provide easy content identification for instruments or color code items, etc.		
Preparation for Use	Place new roll of tape in the dispenser.		
Diagrams (drawings, pictures):			
Steps for Use of Product	 Use tape to close packages and items for sterilization. Indicator (Class 1) tapes with STEAM indicator changing color from green to brown/black. (The ITSG indicator tapes' color change may vary due to differences in processing parameters.) Indicator tapes with EtO gas indicator changing color from brown to green. 		
Interpretation of Results			
Contraindications of Test Results			
Documentation			
Special Warnings and Cautions	 Tape is single-use only The indicators should not be exposed to direct light prior to use. 		
Disposal			

Reprocessing Instructions		
Point of use:		
Preparation for decontamination:		
Disassembly Instructions:		
Cleaning – Manual:		
Cleaning - Automated:		
Disinfection:		
Drying:		
Maintenance, inspection, and testing:		

Reassembly Instructions:	
Packaging:	
Sterilization:	Steam The steam indicator ink will react positively when subjected to heat and moisture in the steam autoclave. After sterilization strong dark lines appear on the backing. • Pre-vac at 4 minutes/270° F, 3 minutes at 275° F • Gravity at 15 minutes at 250° F. EO The ethylene oxide indicator ink will react positively when subjected to ethylene oxide autoclave process.
Storage:	
Additional Information:	

Related Healthmark Products	
Other Product Support Documents	Labeling Brochure, Labeling Price List
Reference Documents	
Customer Service contact:	Healthmark Industries Company, Inc
	33671 Doreka
	Fraser, MI 48026
	1-586-774-7600
	healthmark@hmark.com
	hmark.com

2015-11-30 Ralph J Basile