



**Instructions for Use: HTK Labels**

<b>Brand Name of Product</b>	HangTimeKit™ Labels
<b>Generic Name of Product</b>	Kit for managing and insuring the safe storage of flexible endoscopes
<b>Product Code Number(s)</b>	403225 HANG TIME, 403225 HTKG, 403225 HTKY, 403225 HTKB, 403225 HTKO, 403225 HTKP
<b>Purpose of Product</b>	The HangTime label is for documenting the date of reprocessing.
<b>Range of Applications for Product</b>	Surveillance of flexible endoscopes in storage to track the date of when the endoscope was last reprocessed.
<b>Key specifications of product</b>	<ul style="list-style-type: none"> <li>Numbers representing the days of the month and abbreviations for the 12 months are printed on the label.</li> <li>Offered in a variety of colors.</li> </ul>

<b>Shipping &amp; Storage</b>	
<b>Shipping Conditions &amp; Requirements</b>	
<b>Storage Conditions</b>	
<b>Packaging Conditions</b>	
<b>Shelf Life</b>	

<b>Instructions for Using Product</b>	
<b>Description of Use (s)</b>	The label is for recording and documenting the latest date that the flexible endoscope was reprocessed.
<b>Preparation for Use</b>	
<b>Diagrams (drawings, pictures):</b>	
<b>Steps for Use of Product</b>	<ol style="list-style-type: none"> <li>Circle the day and month of the last reprocessed date.</li> <li>Optionally include the initials of the reprocessor.</li> <li>Place the label securely around the scope through the looped opening.</li> </ol>
<b>Interpretation of Results</b>	If the endoscope has been in storage for an extended amount of time, you may need to test it with the HangTimeKit™ to be sure it's still safe to use.
<b>Contraindications of Test Results</b>	
<b>Documentation</b>	
<b>Special Warnings and Cautions</b>	
<b>Disposal</b>	Labels are single use only. When finished, it should be treated as biohazard and disposed of in accordance to your facilities guidelines.

<b>Reprocessing Instructions</b>	
<b>Point of use:</b>	
<b>Preparation for decontamination:</b>	
<b>Disassembly Instructions:</b>	
<b>Cleaning – Manual:</b>	
<b>Cleaning – Automated:</b>	
<b>Disinfection:</b>	
<b>Drying:</b>	
<b>Maintenance, inspection, and testing:</b>	
<b>Reassembly Instructions:</b>	
<b>Packaging:</b>	
<b>Sterilization:</b>	
<b>Storage:</b>	
<b>Additional Information:</b>	

<b>Related Healthmark Products</b>	HangTimeKit™
<b>Other Product Support Documents</b>	Hang Time Kit Specification Sheet, HangTimeKit Pocket Brochure, HangTimeKit Whitepaper, HTK-FV1041 Gram Negative Test Specification Sheet, Illustrated Users Guide for HangTime test
<b>Reference Documents</b>	
<b>Customer Service contact:</b>	Healthmark Industries Company, Inc

	33671 Doreka Fraser, MI 48026 1-586-774-7600 <a href="mailto:healthmark@hmark.com">healthmark@hmark.com</a> hmark.com
--	---

2015-11-13 Ralph J Basile