

Instructions for Use: HTK Labels

Brand Name of Product	HangTimeKit™ Labels
Generic Name of Product	Kit for managing and insuring the safe storage of flexible endoscopes
Product Code Number(s)	403225 HANG TIME, 403225 HTKG, 403225 HTKY, 403225 HTKB, 403225 HTKO, 403225 HTKP
Purpose of Product	The HangTime label is for documenting the date of reprocessing.
Range of Applications for Product	Surveillance of flexible endoscopes in storage to track the date of when the endoscope was last reprocessed.
Key Specifications of Product	<ul style="list-style-type: none"> Numbers representing the days of the month and abbreviations for the 12 months are printed on the label. Offered in a variety of colors.

Shipping & Storage

Shipping Conditions & Requirements	
Storage Conditions	
Packaging Conditions	
Shelf Life	

Instructions for Using Product

Description of Use (s)	The label is for recording and documenting the latest date that the flexible endoscope was reprocessed.
Preparation for Use	
Diagrams (drawings, pictures):	
Steps for Use of Product	<ol style="list-style-type: none"> Circle the day and month of the last reprocessed date. Optionally include the initials of the reprocessor. Place the label securely around the scope through the looped opening.
Interpretation of Results	If the endoscope has been in storage for an extended amount of time, you may need to test it with the HangTimeKit™ to be sure it's still safe to use.
Contraindications of Test Results	
Documentation	
Special Warnings and Cautions	
Disposal	Labels are single use only. When finished, it should be treated as biohazard and disposed of in accordance to your facilities guidelines.

Reprocessing Instructions

Point of Use	
Preparation for Decontamination	
Disassembly Instructions	
Cleaning – Manual	
Cleaning – Automated	
Disinfection	
Drying	
Maintenance, Inspection, and Testing	
Reassembly Instructions	
Packaging	
Sterilization	
Storage	
Additional Information	

Related Healthmark Products	HangTimeKit™
Other Product Support Documents	Hang Time Kit Specification Sheet, HangTimeKit Pocket Brochure, HangTimeKit Whitepaper, HTK-FV1041 Gram Negative Test Specification Sheet, Illustrated Users Guide for HangTime test
Reference Documents	
Customer Service Contact	Healthmark Industries Company, Inc

	18600 Malyn Blvd. Fraser, MI 48026 1-586-774-7600 healthmark@hmark.com hmark.com
--	--

2021-02-09 Ralph J Basile