



Instructions for Use: Instrument Marking Tape

Brand Name of Product	Instrument Marking Tape
Generic Name of Product	Instrument Marking Tape
Product Code Number(s)	MKT001 – MKT033
Purpose of Product	Marking tapes which allow the effective marking of instruments to aide their proper processing and packaging. .
Range of Applications for Product	Surgical instruments and other medical devices requiring distinctive identification
Key specifications of product	<ul style="list-style-type: none"> • Tape material - 66µ Calendared, cadmium-free polyvinyl chloride (PVC) based film <ul style="list-style-type: none"> ○ Adhesive - Permanent, acrylic adhesive ○ Backing - 135 g/m2 silicone, white clay-coated Kraft paper ○ Marking Tape is Latex-Free • Tensile strength >1.5 kN/m • Adhesive strength against steel >15 N/25mm • Solvent and Chemical Resistance <ul style="list-style-type: none"> ○ Resistant to most oils, greases and fuel • Aliphatic solvents, weak acids and alkalis

Shipping & Storage

Shipping Conditions & Requirements	
Storage Conditions	
Packaging Conditions	
Shelf Life	

Instructions for Using Product

Description of Use (s)	
Preparation for Use	
Diagrams (drawings, pictures):	
Steps for Use of Product	<ol style="list-style-type: none"> 1. Peel adhesive label off paper. The tape is most easily removed from the sheet by folding the sheet along the score line and lifting from one end. 2. The tape should be applied on a part of the instrument where it is not tapered and therefore will remain flat. Do not apply where it may interfere with the performance of the instrument ore where it may affect the handling. 3. Secure adhesive back to instrument being marked. The tape wrap around itself at least once. 4. The instrument should then be autoclaved. This effectively shrinks the tape around the instrument and sets the tape to tape adhesion. 5. The instrument can then be washed in the normal fashion and then be circulated. During normal usage, the tape should retain it color for about 2-6 months. 6. It may be removed by cutting with a sharp knife. Any adhesive remaining on the instrument may be removed by scrubbing with a scouring pad. Hot wated to soften the adhesive may help with the removal.
Interpretation of Results	
Contraindications of Test Results	
Documentation	
Special Warnings and Cautions	
Disposal	

Reprocessing Instructions

Point of use:	
Preparation for decontamination:	
Disassembly Instructions:	
Cleaning – Manual:	
Cleaning – Automated:	
Disinfection:	
Drying:	

Maintenance, inspection, and testing:	
Reassembly Instructions:	
Packaging:	
Sterilization:	<ul style="list-style-type: none"> • Compatible with steam sterilization. • Not compatible with EO sterilization • Not compatible with gas-plasma sterilization
Storage:	
Additional Information:	Marking Tape complies with the essential requirements of the Medical Devices Directive EC 93/42 -2007/47/EC - Class 1. It is CE – marked on the product label

Related Healthmark Products	
Other Product Support Documents	Labeling Price List, Labeling Brochure
Reference Documents	
Customer Service contact:	Healthmark Industries Company, Inc 33671 Doreka Fraser, MI 48026 1-586-774-7600 healthmark@hmark.com hmark.com

2014-08-07 Ralph J Basile