




**Instructions for Use: Interchangeable Holder & Tags**

<b>Brand Name of Product</b>	Interchangeable Label Holder & Tags
<b>Generic Name of Product</b>	Labels
<b>Product Code Number(s)</b>	01498, 01948, 01949, 01950, 01969, 01951, 01952, 01953, 01954, 10990, 01991
<b>Intended Use</b>	Labels are manufactured with a durable clip that attaches to various medical trays and mesh wire baskets. Accompanied by interchangeable color tags for adequate identification, these can be laser imprinted with text and barcode (green labels can only be engraved with text).
<b>Range of Applications for Product</b>	Mesh Basket Instrument Sterilization Containers
<b>Key specifications of product</b>	<ul style="list-style-type: none"> <li>• Clips on various metal trays</li> <li>• 60 x 40 mm Label holder is made from Stainless Steel</li> <li>• Color Tags are made from Polyethersulfone</li> <li>• Heat resistant up to 135 °C or 275 °F</li> <li>• Color tags are available in variety of colors</li> </ul>

**Shipping & Storage**

<b>Shipping Conditions &amp; Requirements</b>	
<b>Storage Conditions</b>	
<b>Packaging Conditions</b>	
<b>Shelf Life</b>	

**Instructions for Using Product**

<b>Description of Use (s)</b>	The stainless steel label holder is suitable for sterilization and are resistant to washing and disinfection and thus can stay with a tray of instruments all the way through the decontamination process.
<b>Preparation for Use</b>	
<b>Diagrams (drawings, pictures):</b>	
<b>Steps for Use of Product</b>	<ol style="list-style-type: none"> <li>1. Attach the clips of the label to the metal tray or wire basket.</li> <li>2. Insert color tag into the holder.</li> </ol>
<b>Interpretation of Results</b>	
<b>Contraindications of Test Results</b>	
<b>Documentation</b>	
<b>Special Warnings and Cautions</b>	<ul style="list-style-type: none"> <li>• Do not handle basket labels until they are allowed to cool down to room temperature after sterilization or thermal disinfection.</li> </ul>
<b>Disposal</b>	

**Reprocessing Instructions**

<b>Point of use:</b>	
<b>Preparation for decontamination:</b>	

<b>Disassembly Instructions:</b>	
<b>Cleaning – Manual:</b>	
<b>Cleaning – Automated:</b>	
<b>Disinfection:</b>	
<b>Drying:</b>	
<b>Maintenance, inspection, and testing:</b>	
<b>Reassembly Instructions:</b>	
<b>Packaging:</b>	
<b>Sterilization:</b>	<ul style="list-style-type: none"> <li>• Compatible with steam, EO and gas-plasma/VHP sterilization.</li> </ul>
<b>Storage:</b>	
<b>Additional Information:</b>	

<b>Related Healthmark Products</b>	
<b>Other Product Support Documents</b>	Label Price List, Label Brochure
<b>Reference Documents</b>	
<b>Customer Service contact:</b>	Healthmark Industries Company, Inc 33671 Doreka Fraser, MI 48026 1-586-774-7600 <a href="mailto:healthmark@hmark.com">healthmark@hmark.com</a> <a href="http://hmark.com">hmark.com</a>

2016-08-01 Ashlynn Basile