



04/05/2010

**CLINIPAK TRAY TAGS – COMPATIBLE STERILISATION METHODS**

I hereby confirm that Clinipak Tray Tags with the Following Product Codes are compatible with Steam (up to 137°C), Ethylene Oxide and Plasma Sterilisation Methods:

TAG001 – TAG026 inclusive  
TAG55WH

A handwritten signature in black ink that reads "Andy Wigglesworth". The signature is written in a cursive style with a horizontal line underneath.

Andy Wigglesworth  
QA Manager  
Clinipak Ltd

Tel: 01628 648091  
Email: [qualitycontrol@clinipak.co.uk](mailto:qualitycontrol@clinipak.co.uk)  
Fax 01628 810760

Clinipak Limited · Beech House · Knaves Beech Business Centre · Loudwater · High Wycombe · Bucks · HP10 9SD  
Tel: +44 (0) 1628 810626 Fax: +44 (0) 1628 810760 Email: [sales@clinipak.co.uk](mailto:sales@clinipak.co.uk) [www.clinipak.co.uk](http://www.clinipak.co.uk)