



Instructions for Use: Tray Identification Tags

Brand Name of Product	Tray Identification Tags
Generic Name of Product	Tray Identification Tags
Product Code Number(s)	RIN001, TG001, TG001-E, TG011, TG012, TG013, TG014, TG015, TG016, TG017, TG018, TG019, TG020, TG021, TG022, TG023, TG024, TG025, TG026, TG011-E, TG012-E, TG013-E, TG014-E, TG015-E, TG016-E, TG017-E, TG018-E, TG019-E, TG020-E, TG021-E, TG022-E, TG023-E, TG024-E, TGO25-E, TG026-E, TIE001, TIE002, TAG-CONTAIN2
Intended Use	Tray Tags are a range of plastic labeling tags, which are intended to be used for tracking and tracing purposes.
Range of Applications for Product	For use with trays, baskets and other items for identification purposes.
Key Specifications of Product	<ul style="list-style-type: none"> • Resin bonded • Various colors • X-ray detectable • Autoclavable

Shipping & Storage	
Shipping Conditions & Requirements	N/A
Storage Conditions	N/A
Packaging Contents	N/A
Shelf Life	N/A

Instructions for Using Product	
Description of Use(s)	For use with trays, baskets, and other items for identification purposes.
Preparation for Use	N/A
Diagrams (drawings, pictures):	N/A
Steps for Use of Product	<p>HOW TO USE TRAY TAGS</p> <p>Writing directly onto the Tray Tag:</p> <ol style="list-style-type: none"> 1. Information may be placed onto a Tray Tag by writing directly on it using a suitable marking pen. Care should be taken to ensure the writing will not be removed during reprocessing. Recommendation: test this prior to placing in use. 2. Either or both sides of the Tag may be written on: <div data-bbox="522 1444 1539 1692" data-label="Image"> </div> <p>Using a pre-printed self-adhesive label:</p> <p>Pre-printed self-adhesive labels may be stuck onto either side of a Tag. Care should be taken to ensure the label and its information are compatible with reprocessing. It is recommended to test this first prior to placing in use.</p>



When Attaching the Tray Tag to Tray or Basket or other items:

1. Use the holes in the Tray Tag to attach it to the tray or basket.
2. Use a split ring (RIN001) or a Security Tie (TIE001 or TIE002) for this purpose.
3. If X-Ray detectability is required, use a split ring, RIN001, to attach the Tray Tag to the tray since the split ring is X-Ray detectable.



Interpretation of Results	N/A
Contraindications of Test Results	N/A
Documentation	N/A
Special Warnings and Cautions	<ul style="list-style-type: none"> ● The facility should develop a policy for the placement of the identification tag on the tray or basket. This should include a risk assessment to avoid potential interference with effective processing or aseptic presentation. ● Tray Tags should be stored away from any sources of ignition.
Disposal	<ul style="list-style-type: none"> ● Unused Tray Tags may be disposed of in the plastic waste stream. ● Used Tray Tags should be disposed of by considering possible microbial contamination. ● Users should reference facility procedures.

Reprocessing Instructions	
Point of Use	N/A
Preparation for Decontamination	N/A
Disassembly Instructions	N/A
Cleaning – Manual	N/A
Cleaning – Automated	Tray Tags are compatible with washer-disinfectors conforming to ISO 15883 series of standards.
Disinfection	N/A
Drying	N/A

Maintenance, Inspection, and Testing	N/A
Reassembly Instructions	N/A
Packaging	N/A
Sterilization	<ul style="list-style-type: none"> ● Materially compatible with Steam, EO and VHP Sterilization ● Recommended cycle for steam sterilization: <ul style="list-style-type: none"> ○ 132 °C (270 °F) for 4 minutes ○ 135 °C (275 °F) for 3 minutes.
Storage	N/A
Additional Information	N/A

Related Healthmark Products	N/A
Other Product Support Documents	Labeling Brochure, Labeling Price List
Reference Documents	N/A
Customer Service Contact	Healthmark Industries Company, Inc. 18600 Malyn Blvd. Fraser, MI 48026 1-586-774-7600 healthmark@hmark.com hmark.com