



**Instructions for Use: Self-Looping labels IFU**

<b>Brand Name of Product</b>	Self-Looping Labels
<b>Generic Name of Product</b>	Tray labels
<b>Product Code Number(s)</b>	403005, 403115, 403125, 403225, 404225-351-GI, 403225 LOANER, 403225 PINK, 403225 RED, 403225 RED BIO, 403225 SLOTTED, 403225 SLOTTED PINK, 403225 YELLOW, 403225 BLUE, 403225 GREEN, TAG-CONTAIN2
<b>Purpose of Product</b>	These labels facilitate specific markings and remarks for the sterile contaminate trays. They are designed to minimize misunderstandings about the contents of an instrument tray and to notify about maintenance of an instrument.
<b>Range of Applications for Product</b>	Instrument Sterilization Containers
<b>Key specifications of product</b>	<ul style="list-style-type: none"> <li>• Self-looping style for mounting on the handle of any closed container</li> <li>• Compatible with steam, EO and gas-plasma sterilization</li> <li>• Available in a variety of colors</li> </ul>

**Shipping & Storage**

<b>Shipping Conditions &amp; Requirements</b>	
<b>Storage Conditions</b>	
<b>Packaging Conditions</b>	Packaged 500 per box
<b>Shelf Life</b>	

**Instructions for Using Product**

<b>Description of Use (s)</b>	The PET- based label is suitable for sterilization and are resistant to washing and disinfection and thus can stay with a tray of instruments all the way through the decontamination process.
<b>Preparation for Use</b>	
<b>Diagrams (drawings, pictures):</b>	
<b>Steps for Use of Product</b>	<ol style="list-style-type: none"> <li>1. You will loop the tray label through the handle of the container or tray.</li> <li>2. Push the count sheet through the tray label so it is evenly hanging on the slotted tray label. You can also affix an ID label to the front of the tag.</li> </ol>
<b>Interpretation of Results</b>	
<b>Contraindications of Test Results</b>	
<b>Documentation</b>	
<b>Special Warnings and Cautions</b>	<ul style="list-style-type: none"> <li>• Tray labels should not be in physical contact with surgical instruments during sterilization cycles.</li> <li>• Do not handle tray labels until they are allowed to cool down to room temperature after sterilization or thermal disinfection.</li> </ul>
<b>Disposal</b>	

**Reprocessing Instructions**

<b>Point of use:</b>	
<b>Preparation for decontamination:</b>	
<b>Disassembly Instructions:</b>	
<b>Cleaning – Manual:</b>	
<b>Cleaning – Automated:</b>	
<b>Disinfection:</b>	
<b>Drying:</b>	
<b>Maintenance, inspection, and testing:</b>	
<b>Reassembly Instructions:</b>	
<b>Packaging:</b>	
<b>Sterilization:</b>	<ul style="list-style-type: none"> <li>• Compatible with steam, EO and gas-plasma/VHP sterilization.</li> <li>• Recommended cycle time for steam sterilization is 270°F – for 4 minutes / 275°F for 3 minutes</li> </ul>
<b>Storage:</b>	
<b>Additional Information:</b>	

<b>Related Healthmark Products</b>	
<b>Other Product Support Documents</b>	Labels Brochure, Labels Price List
<b>Reference Documents</b>	
<b>Customer Service contact:</b>	Healthmark Industries Company, Inc 33671 Doreka Fraser, MI 48026 1-586-774-7600 <a href="mailto:healthmark@hmark.com">healthmark@hmark.com</a> <a href="http://hmark.com">hmark.com</a>

2016-08-01 msmith