

<b>Brand Name of Product</b>	Cable-Tie Style Loks
<b>Generic Name of Product</b>	Tamper Evident Seal
<b>Product Code Number(s)</b>	8525 WHITE
<b>Purpose of Product</b>	To show tampering and/or access to critical equipment and supplies.
<b>Range of Applications for Product</b>	Crash carts, emergency boxes, cabinets, and other areas where demonstrating tampering is critical.
<b>Key Specifications of Product</b>	<ul style="list-style-type: none"> <li>• Numbered</li> <li>• Cynch-tight nylon security straps with a steel locking tooth</li> <li>• Seal and broken by hand</li> <li>• Cannot be resealed once broken</li> </ul>

**Shipping & Storage**

<b>Shipping Conditions &amp; Requirements</b>	
<b>Storage Conditions</b>	
<b>Packaging Contents</b>	
<b>Shelf Life</b>	

**Instructions for Using Product**

<b>Description of Use (s)</b>	Seal crash carts, emergency boxes and other areas where evidence of tampering is critical when a longer seal is needed
<b>Preparation for Use</b>	
<b>Diagrams (drawings, pictures)</b>	
<b>Steps for Use of Product</b>	<ol style="list-style-type: none"> <li>1. Insert seal into closing of two sides of the object being secured.</li> <li>2. Seal through the locking dome of the seal.</li> <li>3. Pull tail of seal as tight as possible to ensure item is tightly sealed.</li> <li>4. The seal can be easily opened with a tug.</li> </ol>
<b>Interpretation of Results</b>	
<b>Contraindications of Test Results</b>	
<b>Documentation</b>	Follow the policy of your facility for recording date, staff and other information critical to management of critical inventory.
<b>Special Warnings and Cautions</b>	If a seal is broken, assume items have been tampered with or removed. Follow your facility instructions for inspection and replacement of critical inventory.
<b>Disposal</b>	

**Reprocessing Instructions**

<b>Point of Use</b>	
<b>Preparation for Decontamination</b>	
<b>Disassembly Instructions</b>	
<b>Cleaning – Manual</b>	
<b>Cleaning – Automated</b>	
<b>Disinfection</b>	
<b>Drying</b>	
<b>Maintenance, Inspection, and Testing</b>	
<b>Reassembly Instructions</b>	
<b>Packaging</b>	
<b>Sterilization</b>	<ul style="list-style-type: none"> <li>• Compatible with steam sterilization</li> <li>• Standard cycle: 132°C at 4 minutes, 135°C at 3 minutes</li> </ul>
<b>Storage</b>	
<b>Additional Information</b>	

<b>Related Healthmark Products</b>	
<b>Other Product Support Documents</b>	Security Products Brochure, Security Products Price List
<b>Reference Documents</b>	
<b>Customer Service Contact</b>	Healthmark Industries Company, Inc

	18600 Malyn Blvd. Fraser, MI 48026 1-586-774-7600 <a href="mailto:healthmark@hmark.com">healthmark@hmark.com</a> <a href="http://hmark.com">hmark.com</a>
--	---

2020-08-28 msmith