



**Instructions for Use: Plastic & Wire Loks**

<b>Brand Name of Product</b>	Plastic & Wire Loks
<b>Generic Name of Product</b>	Tamper Evident Seals
<b>Product Code Number(s)</b>	PWR-6
<b>Purpose of Product</b>	To show tampering and/or access to critical equipment and supplies.
<b>Range of Applications for Product</b>	Crash carts, emergency boxes, cabinets, and other areas where demonstrating tampering is critical.
<b>Key Specifications of Product</b>	<ul style="list-style-type: none"> <li>• Numbered</li> <li>• Seal and broken by hand</li> <li>• Cannot be resealed once broken</li> <li>• Variety of colors</li> <li>• Manufactured of tough plastic body and metal wire hasp</li> <li>• Write-on panel on one or two sides</li> </ul>

**Shipping & Storage**

<b>Shipping Conditions &amp; Requirements</b>	
<b>Storage Conditions</b>	
<b>Packaging Contents</b>	100 per box
<b>Shelf Life</b>	

**Instructions for Using Product**

<b>Description of Use (s)</b>	Seal crash carts, emergency boxes and other areas where evidence of tampering is critical
<b>Preparation for Use</b>	
<b>Diagrams (drawings, pictures):</b>	
<b>Steps for Use of Product</b>	<ol style="list-style-type: none"> <li>1. Pull on wire to remove loose hasp end from seal body</li> <li>2. Pass open end through opening of object to be sealed</li> <li>3. Re-insert hasp end into slot seal body. Firmly push down hasp until it is securely sealed. Pull on seal to ensure locking mechanism is engaged.</li> <li>4. Twist seal body to break and remove.</li> </ol>
<b>Interpretation of Results</b>	
<b>Contraindications of Test Results</b>	
<b>Documentation</b>	Follow the policy of your facility for recording date, staff and other information critical to management of critical inventory.
<b>Special Warnings and Cautions</b>	If a seal is broken, assume items have been tampered with or removed. Follow your facility instructions for inspection and replacement of critical inventory.
<b>Disposal</b>	

**Reprocessing Instructions**

<b>Point of Use</b>	
<b>Preparation for Decontamination</b>	
<b>Disassembly Instructions</b>	
<b>Cleaning – Manual</b>	
<b>Cleaning – Automated</b>	
<b>Disinfection</b>	
<b>Drying</b>	
<b>Maintenance, Inspection, and Testing</b>	
<b>Reassembly Instructions</b>	
<b>Packaging</b>	
<b>Sterilization</b>	

<b>Storage</b>	
<b>Additional Information</b>	
<b>Related Healthmark Products</b>	
<b>Other Product Support Documents</b>	Security Products Brochure, Security Products Price List
<b>Reference Documents</b>	
<b>Customer Service Contact</b>	Healthmark Industries Company, Inc. 18600 Malyn Blvd. Fraser, MI 48026 1-586-774-7600 <a href="mailto:healthmark@hmark.com">healthmark@hmark.com</a> <a href="http://hmark.com">hmark.com</a>