



**Instructions for Use: Container Loks**

<b>Brand Name of Product</b>	Container Loks
<b>Generic Name of Product</b>	Closed sterilization container tamper evident seal
<b>Product Code Number(s)</b>	3030 BL, 3030 BLDTS, 3030 BLWDTS, 3030 WT, 3030 WTDTS, 3030 YL, 3030 YLDTS, 4040 GN, 4040 GNDTS, 5030 RDDTS, 6020 WTDTS, 9393 RDWDTS, 9393 WTDTS & 605M GRWDTS, 605LNG GR, 605M GR, 605M GRDT, 201226 WDTs.
<b>Purpose of Product</b>	Seal closed sterilization container with a tamper evident lock to insure container has not been open since sterilization.
<b>Range of Applications for Product</b>	
<b>Key Specifications of Product</b>	<ul style="list-style-type: none"> <li>• Blue, Yellow, White, Red, Green (4040 only available in green)</li> <li>• Optional steam sterilization indicator dot (3030YLFD is lead free)</li> </ul>

<b>Shipping &amp; Storage</b>	
<b>Shipping Conditions &amp; Requirements</b>	
<b>Storage Conditions</b>	
<b>Packaging Conditions</b>	
<b>Shelf Life</b>	

<b>Instructions for Using Product</b>	
<b>Description of Use (s)</b>	Seal closed sterilization container with a tamper evident lock to insure container has not been open since sterilization.
<b>Preparation for Use</b>	
<b>Diagrams (drawings, pictures)</b>	
<b>Steps for Use of Product</b>	<ol style="list-style-type: none"> <li>1. Pass arrow shackle through opening of item to be sealed.</li> <li>2. Firmly push arrow shackle into locking device until it engages.</li> <li>3. Pull on seal to insure it has been properly locked.</li> </ol>
<b>Interpretation of Results</b>	<ul style="list-style-type: none"> <li>• Steam sterilization: If purchased with optional steam indicator dot, reactive ink will turn brown/dark brown in the presence of steam sterilization (270°F +/- 5°F).</li> </ul>
<b>Contraindications of Test Results</b>	
<b>Documentation</b>	
<b>Special Warnings and Cautions</b>	If a seal is broken, assume sterility has been compromised and the tray must be reprocessed.
<b>Disposal</b>	

<b>Reprocessing Instructions</b>	
<b>Point of Use</b>	
<b>Preparation for Decontamination</b>	
<b>Disassembly Instructions</b>	
<b>Cleaning – Manual</b>	
<b>Cleaning – Automated</b>	
<b>Disinfection</b>	
<b>Drying</b>	
<b>Maintenance, Inspection, and Testing</b>	
<b>Reassembly Instructions</b>	

<b>Packaging</b>	
<b>Sterilization</b>	<ul style="list-style-type: none"> <li>Seal is materially compatible with steam, EO and gas-plasma/VHP sterilization</li> <li>If purchased with optional steam indicator dot: Class 1 chemical indicator. 270°F for up to 4 minutes.</li> </ul>
<b>Storage</b>	<ul style="list-style-type: none"> <li>Should be stored in closed cabinets or otherwise protected from the UV-light source. The process indicator is sensitive to UV-light and may change color over time, if not stored properly.</li> </ul>
<b>Additional Information</b>	

<b>Related Healthmark Products</b>	
<b>Other Product Support Documents</b>	Sterilization Packaging Products Brochure, Sterilization Packaging Products Price List
<b>Reference Documents</b>	
<b>Customer Service Contact</b>	Healthmark Industries Company, Inc 33671 Doreka Fraser, MI 48026 1-586-774-7600 <a href="mailto:healthmark@hmark.com">healthmark@hmark.com</a> hmark.com

2018-06-28 msmith