

Instructions for Use: Pouch Sterilization Rack

Brand Name of Product	Pouch Sterilization Rack
Generic Name of Product	Sterilization Rack
Product Code Number(s)	SR-048, SR-010, SR-040, SR-004, SR-025, SR-005, SR-008, SR-016, SR-020
Purpose of Product	Keeps pouches on edge preventing over stacking, facilitates sterilant contact and reduces drying time. All in accordance with AAMI ST79
Range of Applications for Product	
Key specifications of product	Constructed of stainless steel rod and tungsten inert gas welded for extra strength. Styles include: <ul style="list-style-type: none"> • Full Loop • Peg • Multi-Height Racks

Shipping & Storage

Shipping Conditions & Requirements	
Storage Conditions	
Packaging Conditions	
Shelf Life	

Instructions for Using Product

Description of Use (s)	SRS's peel pouch racks are designed for optimal airflow for the sterilization cycle and allow for complete drying of peel pouches. We offer an entire line of peel pack sizes and configurations customized for your requirements.
Preparation for Use	
Diagrams (drawings, pictures):	
Steps for Use of Product	<ol style="list-style-type: none"> 1. Place sterilization rack in the designated area of the sterilizer. 2. Place each sealed pouch in each slot so there is space between each pouch. 3. When sterilization is complete, remove each pouch from the rack.
Interpretation of Results	
Contraindications of Test Results	
Documentation	
Special Warnings and Cautions	
Disposal	

Reprocessing Instructions

Point of use:	
Preparation for decontamination:	
Disassembly Instructions:	
Cleaning – Manual:	Wash with a mild detergent. Rinse thoroughly.
Cleaning – Automated:	
Disinfection:	
Drying:	
Maintenance, inspection, and testing:	
Reassembly Instructions:	
Packaging:	
Sterilization:	Compatible with steam, EO and gas plasma/VHP sterilization.
Storage:	
Additional Information:	

Related Healthmark Products	
Other Product Support Documents	Sterilization Brochure, Sterilization Price List
Reference Documents	
Customer Service contact:	Healthmark Industries Company, Inc 33671 Doreka Fraser, MI 48026

	1-586-774-7600 healthmark@hmark.com hmark.com
--	---

2016-08-01 Matt Smith